

Certificate



Certificate no.

CU 72160034 01

License Holder:

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3
Canada

Manufacturing Plant:

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3
Canada

Test report no.: USA-EF 31690047 002

Client Reference: Mike Jameson

Tested to: ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 NO. 60601-1-08 + A1

Certified Product: Portable Tourniquet System

License Fee - Units

Model Designation: PTSii (9-2200-001, 60-1000-101-00),
PTS for BFR (9-2200-001BFR)

7

Rated Voltage: AC 100-240V, 50/60Hz (external power supply)
Rated Current: 2.3A (external power supply)
Protection Class: I (external power supply)
Output Ratings DC: 12V, 8.3A (external power supply)

Appendix: 1, 1-2

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

09/02/2016



Business Stream Products
Certification Services

Date : 2016/02/12

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3
Canada

Attn: Mike Jameson

Re. : CU US + Canada Certificate

Type of Equipment : Portable Tourniquet System
Model Designation : See Certificate
Certificate No. : CU 72160034 0001
File No. : 31690047 002
Engineer/Contact : Ernesto Fernandez
Standard(s) : ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 NO. 60601-1-08 + A1

Dear Mr. Jameson,


The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72160034 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. M. Glagla
QA Certification Officer

TUV Rheinland
of North America, Inc.
North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Enclosure

Member of
TUV Rheinland Group