

# Certificate



Certificate no.

CU 72201907 01

**License Holder:**

Delfi Medical Innovations, Inc  
1099 8th Ave, West, Suite 106  
Vancouver BC V6H 1C3  
Canada

**Manufacturing Plant:**

Delfi Medical Innovations, Inc  
1099 8th Ave, West, Suite 106  
Vancouver BC V6H 1C3  
Canada

**Test report no.:** USA-JR 31690047 008

**Client Reference:** Tom Lai

**Tested to:**

ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1  
CAN/CSA-C22.2 No. 60601-1:14  
IEC 60601-1-6:2010+A1  
IEC 60601-1-8:2006+A1  
CAN/CSA-C22.2 NO. 60601-1-6:11 + A1  
CAN/CSA-C22.2 NO. 60601-1-8:08 + A1

**Certified Product:** Portable Tourniquet System

**License Fee - Units**

Model Designation: PTSii (9-2200-001, 60-1000-101-00),  
PTS for BFR (9-2200-001BFR)

7

Rated Voltage: AC 100-240V, 50/60Hz (external power supply)  
Rated Current: 2.3A (external power supply)  
Protection Class: I (external power supply)  
Output Ratings DC: 12V, 8.3A (external power supply)

Special Remarks: Replaces Certificate CU72160034.

7

Appendix: 1, 1-3

**Licensed Test mark:**

**Date of Issue**

(day/mo/yr)

04/06/2020

