



Ref. Certif. No.

US-TUVR-012175

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Portable Tourniquet System

Name and address of the applicant

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Name and address of the manufacturer

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Name and address of the factory

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Ratings and principal characteristics

external power supply: AC 100-240V, 50/60Hz; 2.3A; Class I
Output: DC 12V, 8.3A

Trademark (if any)

Delfi

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PTSii (9-2200-001, 60-1000-101-00),
PTS for BFR (9-2200-001BFR)

Additional information (if necessary may also be reported on page 2)

Replaces CB Certificate US-TUVR-12005.

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1
IEC 60601-1-6:2010+A1
IEC 60601-1-8:2006+A1
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31690047 007

This CB Test Certificate is issued by the National Certification Body



TUV Rheinland of North America, Inc.
12 Commerce Road
Newtown, CT 06470 - USA

Date: 01.06.2020

Signature:

Martin Glagla