

**IECEE**

TM

Ref. Certif. No.

DE 2-044515

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME****CB TEST CERTIFICATE**

Product

Portable Tourniquet System

Name and address of the applicant

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Name and address of the manufacturer

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Name and address of the factory

Delfi Medical Innovations, Inc
1099 8th Ave, West, Suite 106
Vancouver BC V6H 1C3, Canada

Ratings and principal characteristics

external power supply: AC 100-240V, 50/60Hz; 100W;
Output: DC 12V
Class II

Trademark (if any)

Delfi

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

PTS for BFR Gen 3

Additional information (if necessary may
also be reported on page 2)A sample of the product was tested and
found to be in conformity withIEC 60601-1:2005+A1+A2
IEC 60601-1-6:2010+A1+A2
IEC 60601-1-8:2006+A1+A2
for national differences see test reportAs shown in the Test Report Ref. No. which
forms part of this Certificate

US25Q5RI 200

This CB Test Certificate is issued by the National Certification Body

**TÜVRheinland®**TÜV Rheinland LGA Products GmbH
Tillystr. 2, 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web : www.tuv.com

Date: 2025-07-17

Signature:

Gregory Chase